

	Freestyle Libre 2 & 3	Dexcom G6 & G7	Dexcom Stelo (OTC)	Medtronic Guardian Connect (GC) and Guardian 4 (G4)	Eversense
Approved for:	≥ 4 years old	≥ 2 years old	≥ 18 years old	14 – 75 years old	≥ 18 years old
Pregnancy Approved:	Yes	G7 only	No	No	No
Frequency of measurement	Every minute	Every 5 minutes	Every 15 minutes	Every 5 minutes	Every 5 minutes
Sensor placement	Arm	G6: Abdomen G7: Arm (>7yo) Arm/buttocks (2-7 yo)	Arm	Arm or Abdomen	Arm - implanted
Sensor wear time	14 days (soon to be 15 days)	10 days	15 days	7 days	180 days
Drug interactions	Vitamin C > 500 mg/day	Acetaminophen (> 4000mg/day) and Hydroxyurea	TBD	Acetaminophen, dose dependent	Tetracyclines
Cash Cost	\$65/sensor, or \$75 for 2 with voucher	\$180/sensor	TBD	Varies by insurance	Varies by insurance and procedure
Voucher:	<a href="https://www.freestyle.abott/us-en/myfreestyle.html">https://www.freestyle.abott/us-en/myfreestyle.html</a> 1-855-632-8658	<a href="https://www.dexcom.com/en-us/savings-center-cgm-without-insurance">https://www.dexcom.com/en-us/savings-center-cgm-without-insurance</a>			<a href="https://www.ascensiadiabetes.com/eversense/coverage/insurance-and-cost/">https://www.ascensiadiabetes.com/eversense/coverage/insurance-and-cost/</a>

References:

1. FDA.gov. Do Not Use Smartwatches or Smart Rings to Measure Blood Glucose Levels: FDA Safety Communication. Feb 21, 2024. <https://www.fda.gov/medical-devices/safety-communications/do-not-use-smartwatches-or-smart-rings-measure-blood-glucose-levels-fda-safety-communication>
2. ADA Professional Practice Committee. 7: Diabetes Technology: Standards of Care in Diabetes – 2024. [https://diabetesjournals.org/care/issue/47/Supplement\\_1](https://diabetesjournals.org/care/issue/47/Supplement_1)
3. Stelo by Dexcom First Glucose Biosensor Cleared by FDA as Over the Counter. March 5, 2024. <https://investors.dexcom.com/news/news-details/2024/Stelo-by-Dexcom-First-Glucose-Biosensor-to-be-Cleared-by-FDA-as-Over-the-Counter/default.aspx>
4. Dexcom Stelo. March 11, 2024. <https://www.dexcom.com/stelo>